




INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

(51) International Patent Classification ⁷ : A61B 5/0492, 5/0488		A1	(11) International Publication Number: WO 00/03637
			(43) International Publication Date: 27 January 2000 (27.01.00)
(21) International Application Number: PCT/CA99/00652 (22) International Filing Date: 16 July 1999 (16.07.99) (30) Priority Data: 2,243,382 16 July 1998 (16.07.98) CA (71) Applicant (for all designated States except US): UNIVER- SITE DE MONTREAL [CA/CA]; 2900 Edouard-Montpetit Boulevard, Montreal, Quebec H3C 3J7 (CA). (72) Inventor; and (75) Inventor/Applicant (for US only): SINDERBY, Christer [CA/CA]; 12750, 27th Avenue, Montreal, Quebec H1A 1Z9 (CA). (74) Agents: DUBUC, Jean, H. et al.; Goudreau Gage Dubuc & Martineau Walker, The Stock Exchange Tower, Suite 3400, 800 Place Victoria, P.O. Box 242, Montreal, Quebec H4Z 1E9 (CA).		(81) Designated States: AE, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CU, CZ, DE, DK, EE, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, UA, UG, US, UZ, VN, YU, ZA, ZW, ARIPO patent (GH, GM, KE, LS, MW, SD, SL, SZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG). Published With international search report.	

(54) Title: DISTURBANCE-FREE ELECTROMYOGRAPHIC PROBE



(57) Abstract

A myographic probe for detecting an electrical signal produced by a muscle and for reducing the influence of electrode disturbances. The probe includes electrodes and a disturbance reducing interface covering each electrode thereby segregating the electrodes from the muscle. Electrode disturbances include problems such as those related to the motion of the electrodes, changes in the pressure applied to the electrode, and/or intermittent contact with surrounding tissue. The disturbance reducing interface is ion permeable and is, when dry, less conductive than the electrodes. The disturbance reducing interface may comprise a matrix of permeable material such as a mesh, foam, or other porous materials. The probe may be in the form of a catheter and be advantageously used in a human cavity such as the oesophagus. Another advantage of the invention is the possibility of using electrodes which are different from conventional wound wire electrodes.

FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AL	Albania	ES	Spain	LS	Lesotho	SI	Slovenia
AM	Armenia	FI	Finland	LT	Lithuania	SK	Slovakia
AT	Austria	FR	France	LU	Luxembourg	SN	Senegal
AU	Australia	GA	Gabon	LV	Latvia	SZ	Swaziland
AZ	Azerbaijan	GB	United Kingdom	MC	Monaco	TD	Chad
BA	Bosnia and Herzegovina	GE	Georgia	MD	Republic of Moldova	TG	Togo
BB	Barbados	GH	Ghana	MG	Madagascar	TJ	Tajikistan
BE	Belgium	GN	Guinea	MK	The former Yugoslav Republic of Macedonia	TM	Turkmenistan
BF	Burkina Faso	GR	Greece			TR	Turkey
BG	Bulgaria	HU	Hungary	ML	Mali	TT	Trinidad and Tobago
BJ	Benin	IE	Ireland	MN	Mongolia	UA	Ukraine
BR	Brazil	IL	Israel	MR	Mauritania	UG	Uganda
BY	Belarus	IS	Iceland	MW	Malawi	US	United States of America
CA	Canada	IT	Italy	MX	Mexico	UZ	Uzbekistan
CF	Central African Republic	JP	Japan	NE	Niger	VN	Viet Nam
CG	Congo	KE	Kenya	NL	Netherlands	YU	Yugoslavia
CH	Switzerland	KG	Kyrgyzstan	NO	Norway	ZW	Zimbabwe
CI	Côte d'Ivoire	KP	Democratic People's Republic of Korea	NZ	New Zealand		
CM	Cameroon			PL	Poland		
CN	China	KR	Republic of Korea	PT	Portugal		
CU	Cuba	KZ	Kazakstan	RO	Romania		
CZ	Czech Republic	LC	Saint Lucia	RU	Russian Federation		
DE	Germany	LI	Liechtenstein	SD	Sudan		
DK	Denmark	LK	Sri Lanka	SE	Sweden		
EE	Estonia	LR	Liberia	SG	Singapore		

DISTURBANCE-FREE ELECTROMYOGRAPHIC PROBE

5

BACKGROUND OF THE INVENTION

1. Field of the invention:

10

The present invention relates to an apparatus for reducing disturbances induced in a signal measurement or recording, in particular, but not exclusively, by movement of the electrodes or changes in the pressure applied to the electrodes.

15

2. Brief description of the prior art:

20

Oesophageal recording of diaphragm electromyogram (EMG) has traditionally been problematic due to the low amplitude of the EMG signal relative to the artifactual disturbances such as, in particular, the so-called electrode motion artifacts. At high gain settings, large electrode motion artifacts lead to saturation of the output of the preamplifier, thereby causing a temporary loss of the EMG signal. This problem of the prior art makes EMG recording very difficult during dynamic manoeuvres, such as for example rapid shallow breathing or panting.

25

OBJECTS AND SUMMARY OF THE INVENTION

5 A first object of the present invention is to provide a technology capable of reducing disturbances induced in a measurement or recording by:

- movements of detecting electrodes;
- changes in the pressure applied to these electrodes; or
- 10 - other mechanical influence on the electrodes, generally referred to as motion artifacts..

 Another object of the present invention is to reduce the amplitude of motion artifacts relative to the amplitude of the EMG signal
15 to thereby reduce the possibility for saturation of the preamplifier.

 A third object of the present invention is to overcome the problems of the prior art related to low signal-to-artifact ratio.

20 A further object of the present invention is to improve bipolar electrode measurements of diaphragm electromyogram (EMG).

 In a preferred embodiment of the invention, there is provided a measurement apparatus for detecting an electrical signal
25 produced by a muscle while reducing signal disturbances caused by motion artifacts, the measurement apparatus comprises:

- a) a probe;
- b) at least one electrode mounted on said probe; and
- c) a disturbance reducing interface attached to said probe

and covering said at least one electrode, the interface being ion permeable and segregating said at least one electrode from the muscle.

5 The objects, advantages and other features of the present invention will become more apparent upon reading of the following non restrictive description of preferred embodiments thereof, given by way of example only with reference to the accompanying drawings.

10

BRIEF DESCRIPTION OF THE DRAWINGS

In the appended drawings:

15

Figure 1 is a schematic representation of a set-up of an EMG analysis system;

20

Figure 2 is a side elevation view of the free end section of an oesophageal catheter on which an array of electrodes of the EMG analysis system of Figure 1 is mounted;

25

Figure 3 is a longitudinal, partial cross sectional view of the free end section of the oesophageal catheter of Figure 2, showing an individual matrix of permeable material applied to each separate electrode of the array;

Figure 4 is a longitudinal, partial cross sectional view of the free end section of the oesophageal catheter of Figure 2, showing a

continuous matrix of permeable material applied to and spanning the entire electrode array;

5 Figure 5 is a partial perspective view of the free end section of an oesophagal catheter, showing an array of semicircular electrodes and a continuous matrix of permeable material applied to and spanning the entire array of semicircular electrodes;

10 Figure 6 is a partial perspective view of the free end section of an oesophageal catheter, showing an array of button electrodes which can be circular, square, rectangular, or of any other shape, and a continuous matrix of permeable material applied to and spanning the entire array of button electrodes;

15 Figure 7 is a longitudinal, partial cross sectional view of the free end section of an oesophageal catheter, showing an electrode embedded in the material of the catheter, and a matrix of permeable material applied to the embedded electrode;

20 Figure 8 is a longitudinal, partial cross sectional view of the free end section of an oesophageal catheter, showing a stud electrode and a matrix of permeable material applied to the stud electrode;

25 Figure 9 is a partial perspective view of the free end section of the oesophageal catheter of Figure 7, showing an array of electrodes such as shown in Figure 7, embedded into the material of the catheter;

Figure 10 is a partial perspective view of the free end

section of an oesophageal catheter, showing an array of button electrodes covered by a matrix of permeable material applied to the outer surface of the oesophageal catheter;

5 Figure 11 is a partial perspective view of the free end section of an oesophageal catheter, showing an array of button electrodes as well as an array of grounding electrodes; and

10 Figure 12 is an end cross sectional view of the array of button electrodes of Figure 10 covered by the matrix applied to the outer surface of the oesophageal catheter.

15 DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENTS

20 The present invention relates to a technology capable of reducing disturbances induced in an electrical signal measurement and/or recording by movement of detecting electrodes or changes in the pressure applied to these electrodes. The electrodes are conductive elements used to detect electrical activity. The range of applications of the present invention includes electrical signal measurement and/or recording wherein electrodes are immersed in an electrolyte (so-called

25 wet electrodes). A typical example is the measurement and/or recording of diaphragm electromyogram (EMG), oesophageal peristalsis, or ECG with electrodes positioned on a catheter which in turn is introduced in the oesophagus.

Although the preferred embodiments will be described hereinafter with reference to oesophageal catheters and an application to the measurement of diaphragm electromyogram (EMG), it should be kept in mind that it is within the scope of the present invention to envisage other applications for this technology using other types of catheters or probes.

Referring to Figures 1 and 2, to measure EMG activity of the diaphragm 11 of a human patient 14, an array of electrodes such as 12 are mounted on the free end section 15 of an oesophageal catheter 13, with an inter-electrode distance d (Figure 2). The distance d is adjusted in relation to body size; distance d will be larger for an adult than for an infant. The catheter 13 is a hollow tube having a diameter related to body size; the diameter will be smaller for infants than for adults. The catheter diameter, electrode size as well as the inter-electrode distance d may also vary in relation to the purpose of the catheter use.

As shown in Figure 1, the catheter 13 is introduced into the patient's oesophagus through one nostril or the mouth until the array of electrodes 12 is situated at the level of the gastro-oesophageal junction. Of course, positioning of the electrode array comprising a series of differentially and axially arranged electrode pairs (for example electrode pairs 1-7 of Figure 2) is guided by the electrocardiographic (ECG) recordings and the diaphragm EMG. Alternatively, the electrodes 12 are monopolar electrodes differentiated in a computer, for example computer 19 of Figure 1. When required, ground is obtained through a separate grounding electrode structure 25 (Figure 1).

FIG. 1 is a schematic diagram of the system for measuring diaphragm EMG activity. It shows a human patient 14 with a catheter 13 inserted into the oesophagus. The catheter 13 has an array of electrodes 12 at its free end section 15. The electrodes 12 are connected to a computer 19. The computer 19 is also connected to a grounding electrode structure 25. The catheter 13 is shown passing through the nostril or mouth of the patient 14.

Positioning of an electrode at the oesophageal hiatus (where the oesophagus passes through the diaphragm is guided by visual inspection and/or computer algorithms studying the intensity, shape and polarity of ECG and diaphragm EMG signals. When the electrode is close to the oesophageal hiatus, i.e. next to the heart, ECG signal amplitude is high. If the electrode array is positioned close to the mouth (away from the heart), ECG signals present lower amplitudes at the proximate electrodes, and higher amplitudes at the distal electrodes. If the electrode array is positioned too far in the stomach, ECG has a high amplitude at the proximate electrodes of the array and a low amplitude at the distal electrodes. If the electrode array spans the region of the heart, ECG signals will show a time shift along the electrode array. If the electrodes are positioned away from the heart, ECG signals show no time lag. Diaphragm EMG signals obtained through electrode pairs located above and below the diaphragm have opposite polarities (with no time shift). EMG signals obtained on the same side of the diaphragm show the same polarity (and no time shift). The characteristics described in this paragraph will help the operator to adequately position the array of electrodes.

20

According to a preferred embodiment, an electrode 12 is mounted on the free end section 15 of the catheter 13 by winding stainless steel wire (not shown) around catheter 13. The wound stainless steel wire presents a rough surface smoothed out by solder, which in turn is electroplated with nickel, copper and then gold or silver. Use of other metallic elements such as semicylindrical electrodes 21 (Figure 5), button electrodes 22 and 23 (Figure 6), etc., could be contemplated. The button electrodes can be arranged into a longitudinal linear array (electrodes

25

22), or at least one button electrode (see 23) can be angularly offset from the electrodes 22 about the longitudinal axis of the catheter section 15.

For larger diameter feeding tubes or catheters, electrodes
5 such as electrode 26 in Figure 7 can be embedded into the material 27 of the feeding tube or catheter 28. Figure 9 shows a longitudinal array of electrodes 26 embedded into the material 27 of the free end section of the oesophageal catheter 28. Figure 9 also shows the electric wires such as 30, embedded in the material 27 of the catheter 28, and individually
10 connecting each electrode 26 to the amplifiers 16 of Figure 1. In the example of Figures 7 and 9, the electrodes 26 are oval. The electric wires such as 30 in Figure 9 individually connect each electrode such as 26 with a respective input of the monopolar or differential (depending on the monopolar or differential arrangement of the electrodes 12 or 26)
15 amplifiers 16 (Figure 1). Obviously, these electric wires 30 follow the catheter such as 28 from the respective electrodes such as 26 to the corresponding amplifiers 16; the electric wires 30 can be embedded in the material such as 27 of the catheter such as 28 or passed separately outside (see for example 45 in Figure 10) or inside (see for example 46
20 in Figure 10) the catheter lumen 47 depending on the intended application. The electric wires such as 30 transmitting the EMG signals collected by the various electrodes such as 26 are necessarily electrically insulated from each other and preferably surrounded by a conductive mesh constituting a shield against external disturbances.

25

Referring now to Figure 8, a stud electrode 31 is illustrated. Each stud electrode 31 is mounted in a hole 32 made through the wall of an oesophageal catheter 33.

The electrodes such as 34 in Figure 10 can also be applied by means of glue or any other suitable adhesive material or compound, including double adhesive tape.

5 In the example of Figures 10 and 12, a linear array of oval electrodes 34 is mounted on the outer surface 44 of a catheter 36 comprising two longitudinal lumens 47 and 48. Referring to Figure 12, each electrode 34 is applied to the catheter surface 44. As described in the foregoing description, the electric wires (see 45 and 46) for
10 individually connecting the electrodes 34 to the amplifiers 16 will extend either inside lumen 47 (see 46 in Figure 10), inside lumen 48, outside the catheter 36 (see 45 in Figure 10), or embedded in the material of the catheter 36.

15 Figure 11 is a partial perspective view the free end section of an oesophageal catheter 37, comprising a longitudinal, linear array of button electrodes 38. Figure 11 also shows an example of grounding electrode structure (see 25 in Figure 1). In the example of Figure 11, the grounding electrode structure comprises a helical array of grounding
20 electrodes 39 mounted on the outer surface 40 of the catheter 37. Of course, the array of grounding electrodes 39 is centered on the longitudinal axis of the catheter 37 and presents the general configuration of a cylindrical helix.

25 Pressure sensors, pH sensors, thermistors and other detector devices can be added onto the catheter in accordance with the requirements of the intended application.

Referring back to Figure 1, the group of amplifiers 16 amplifies and band-pass filters each EMG signal. The amplified EMG signals are sampled by a personal computer 19 through respective isolation amplifiers of a unit 18, to form signal segments of fixed duration.

5 Unit 18 supplies electric power to the various electronic components of the amplifiers 16 and isolation amplifiers while ensuring adequate isolation of the patient's body from such power supply. The unit 18 also incorporates bandpass filters included in the respective EMG signal channels to eliminate the effects of aliasing. The successive EMG signal
10 segments are then digitally processed into the personal computer 19 after analog-to-digital conversion thereof. This analog-to-digital conversion is conveniently carried out by an analog-to-digital converter implemented in the personal computer 19. The personal computer 19 includes a monitor 40 and a keyboard 41.

15 It is believed to be within the capacity of those of ordinary skill in the art to construct suitable amplifiers 16 and an adequate isolation amplifiers and power supply unit 18. Accordingly, the amplifiers 16 and the unit 18 will not be further described in the present specification.

20 To eliminate the problems related to motion of the electrode, changes in the pressure applied to the electrode, and/or intermittent contact with surrounding tissue, a motion artifact reducing interface is applied to the electrode surface. The problems listed above can grouped
25 as disturbances; the motion artifact reducing interface may therefore also be referred to as a disturbance reducing interface. The motion artifact reducing interface advantageously consists of a matrix of permeable material comprising, for example, a mesh, foam or other porous material,

e.g. a fine filament matrix of nylon. The principle of operation is that the matrix of permeable material creates an interface that hosts ions and electrodes and prevents direct contact between the metal surface of the electrode and the surrounding body tissue. The type of permeable material and thickness thereof is not crucial for performance as long as it forms an ion saturated interface producing no direct contact between the electrode and body tissue. However, excessive thickness may cause increased distance between the electrode and muscle, which will weaken the signal strength and lower the frequency content of this signal.

As illustrated in Figures 3-8 and 10, the matrix of permeable material is applied to the exposed surface of the electrodes where the ion concentration gradients are largest to reduce mechanically-caused movements of ions.

The matrix can be formed by separate single matrices individually applied to or integrated in the exposed surface of each electrode. For example, each individual matrix can be glued on, or adhere to by other means, the outer surface of the catheter to cover the associated electrode. However no adhesive material may cover the electrode surface.

The matrix can also take the form of a continuous matrix (Figures 4, 5 and 6) or 35 (Figure 10 and 12). For example, the continuous matrix may form a tube that can be pulled over the catheter to cover the entire span of the array of electrodes (Figure 4), 21 (Figure 5), 22 and 23 (Figure 6), and 34 (Figures 10 and 12). In the case

of a continuous matrix spanning the entire electrode array, the conductivity of the material constituting the matrix, when dry, has to present a conductivity lower than the conductivity of the metal forming the electrodes, whereby electrical conduction is carried out across the matrix, i.e., through the electrolyte. These matrices provide a much more stable voltage with a reduction of the so-called electrode motion induced artifacts on the diaphragm EMG signal.

Also, the matrix can either cover the entire circumference of the catheter (see matrices 20 and 17 of Figures 3-6) or a portion of the circumference of the catheter (see matrices 29, 42 and 35 of Figures 7, 8, 10 and 12). Again these matrices can be adhered to the outer surface of the catheter to cover the electrodes; no adhesive material may cover the electrode surface.

Other alternatives (not shown) are (a) to wind or wrap the matrix around the catheter and the electrodes, and (b) to host or embed the electrodes into the matrix.

The electrode structure according to the invention can be applied to measurement of the diaphragm electromyogram (EMG) exclusively or in combination with a device for providing feeding/medication/liquid supply to the patient, and emptying of gastric liquids, common to the treatment of patients in need of ventilatory support. The electrode structure is usable to provide diaphragm EMG signals from a plurality of conductive elements which in turn can be used to:

- monitor diaphragm EMG (frequency, amplitude or power);
- trigger and control gas flow, gas volume or gas pressure delivered by a mechanical lung ventilator; and
- control a closed loop ventilator system that will
5 automatically adjust the level of inspiratory support in proportion to changes in the neuro-ventilatory efficiency such that the neural drive remains stable at a desired target level.

10 The closed loop ventilator system control can further use the intensity of the diaphragm electromyogram (EMG) obtained immediately before inspiratory flow occurs to quantify pre-inspiratory breathing efforts.

The catheter including the array of electrodes is aimed to be
15 disposed of after a single use; however, when desired, conventional sterilization techniques can be applied in view of re-using the catheter. The catheter can stay in the same patient for extensive periods of time; it is therefore important that the electrodes and matrix be made out of a non-allergen material.

20 Retrocardiac recording of electrocardiogram and oesophageal peristalsis are other possible applications.

The electrode structure according to the invention is
25 applicable in all patients on ventilatory support and will enhance the possibility of obtaining spontaneous breathing and of optimizing patient ventilator interaction. There exists also a utility for this electrode structure during anaesthesia for monitoring vital fonctions of the patient. The

electrode structure can be used in connection with all kinds of ventilator systems in intensive care unit settings or other wards where assisted ventilation is required.

5 Although the present invention has been described hereinabove with reference to preferred embodiments thereof, these embodiments can be modified at will, within the scope of the appended claims, without departing from the spirit and nature of the subject invention.

10

WHAT IS CLAIMED IS:

1. A measurement apparatus for detecting an electrical signal produced by a muscle while reducing signal disturbances caused by motion artifacts, the measurement apparatus comprises:
- 5 a) a probe;
- b) at least one electrode mounted on said probe; and
- c) a disturbance reducing interface attached to said probe and covering said at least one electrode, the interface being ion permeable and segregating said at least one electrode from the muscle.
- 10 2. A measurement apparatus as in claim 1, wherein said disturbance reducing interface is divided to individually cover each electrode of said at least one electrode.
- 15 3. A measurement apparatus as in claim 1, wherein said disturbance reducing interface spans two or more electrodes of said at least one electrode.
- 20 4. A measurement apparatus as in claim 3, wherein said disturbance reducing interface, when dry, presents lower conductivity than said at least one electrode.
- 25 5. A measurement apparatus as in any of the preceeding claims, wherein said disturbance reducing interface comprises a matrix of permeable material.
6. A measurement apparatus as in claim 5, wherein said matrix comprises a mesh.

.16

7. A measurement apparatus as in claim 5, wherein said matrix comprises a foam.
- 5 8. A measurement apparatus as in claim 5, wherein said matrix comprises a porous material.
9. A measurement apparatus as in claim 5, wherein said matrix comprises fine filaments of nylon.
- 10 10. A measurement apparatus as in any of the preceeding claims, wherein said probe is mounted on a catheter for insertion in the body of a mammal to achieve localized readings of an electrical signal produced by a muscle.
- 15 11. A measurement apparatus as in claim 10, wherein said catheter comprises an oesophageal catheter.
12. A measurement apparatus as in claims 10 or 11, wherein said disturbance reducing interface covers the circumference of said catheter.
- 20 13. A measurement apparatus as in claims 10, 11, or 12, wherein said at least one electrode comprises a winding of conductive material wire around said catheter.
- 25 14. A measurement apparatus as in claims 10, 11, or 12, wherein the surface area of said at least one electrode is substantially smaller than the surface area of an electrode comprising a winding of conductive material wire around said catheter.

15. A measurement apparatus as in any of claims 1 to 12, or 14, wherein said at least one electrode forms a longitudinal linear array of electrodes.

5 16. A measurement apparatus as in claim 15, wherein said array comprises one electrode of said array angularly offset from the longitudinal axis formed by the remaining electrodes.

10 17. A measurement apparatus as in any of claims 1 to 16, wherein said at least one electrode comprises a stud electrode.



18. A measurement apparatus as in any of claims 1 to 16, wherein said at least one electrode comprises a button electrode.

15 19. A measurement apparatus as in any of claims 1 to 16, wherein said at least one electrode comprises a semi-cylindrical electrode.

20. A measurement apparatus as in any of the preceeding claims further comprising a grounding electrode structure attached to said probe.

20 21. A measurement apparatus as in claim 20, wherein said grounding electrode structure comprises at least one grounding electrode.

25 22. A measurement apparatus as in any of the preceeding claims wherein the measurement apparatus is inserted in a human cavity.



2/4









FIG. 5

FIG. 6





INTERNATIONAL SEARCH REPORT

International Application No

PCT/CA 99/00652

A. CLASSIFICATION OF SUBJECT MATTER
IPC 7 A61B5/0492 A61B5/0488

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

IPC 7 A61B

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	US 4 582 064 A (SORGER) 15 April 1986 (1986-04-15) column 7, line 54 -column 8, line 42	1,2,4,5, 22
A	US 5 230 786 A (SIEMENS) 27 July 1993 (1993-07-27) column 2, line 30 - line 43 -/-	1,2,4, 10,12, 20-22



Further documents are listed in the continuation of box C.



Patent family members are listed in annex.

* Special categories of cited documents :

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier document but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art.

"&" document member of the same patent family

Date of the actual completion of the international search

13 October 1999

Date of mailing of the international search report

28/10/1999

Name and mailing address of the ISA

European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel. (+31-70) 340-2040, Tx. 31 651 epo nl,
Fax: (+31-70) 340-3016

Authorized officer

Lemercier, D

INTERNATIONAL SEARCH REPORT

Inte. onal Application No

PCT/CA 99/00652

C.(Continuation) DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	<p>KINGMA & AL: "Improved Ag/AgCl pressure electrodes"</p> <p>MEDICAL & BIOLOGICAL ENGINEERING & COMPUTING,</p> <p>vol. 21, no. 3, May 1983 (1983-05), pages 351-357, XP002118744</p> <p>GB</p> <p>Paragraph 2 : "Mechanical construction" ; figure 2</p> <p>Paragraph 5 : "Electrical behaviour of the electrodes" ; figure 6</p> <p>-----</p>	<p>1,2,4,</p> <p>10,11,22</p>

INTERNATIONAL SEARCH REPORT

Information on patent family members

International Application No

PCT/CA 99/00652

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 4582064 A	15-04-1986	DE 3232455 A	01-03-1984
		AT 24106 T	15-12-1986
		WO 8604223 A	31-07-1986
		EP 0104493 A	04-04-1984
US 5230786 A	27-07-1993	DE 59106341 D	05-10-1995
		EP 0455072 A	06-11-1991
		JP 4227228 A	17-08-1992